

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lísem 129/2, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1190147

Product: Baby breathing monitor

Type: BM-02

Rating: 2 x battery AA-LR6 (2 x 1,5 V DC)

Ordering firm: JABLOTRON ALARMS a.s.
Pod Skalkou 4567/33, 466 01 Jablonec nad Nisou, Czech Republic

Manufacturer: JABLOTRON ALARMS a.s.
Pod Skalkou 4567/33, 466 01 Jablonec nad Nisou, Czech Republic

Factories: JabloPCB s.r.o.
U Přehrady 5129/67, 466 04 Jablonec nad Nisou - Mšeno nad Nisou, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 910260-01/01 of: 05.03.2019, 910260-01/02 of: 27.02.2019

A sample of the product was found to be in conformity with:
ČSN EN 60601-1ed.2:2007+A1:2014, ČSN EN 60601-1-2ed.3:2016, ČSN EN 60601-1-6ed.3:2010+A1:2015,
ČSN EN 62366:2008+A1:2015, ČSN EN 60601-1-11ed.2:2016

Other data:

The validity of the certificate is limited to: 31.03.2022

06.03.2019

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



910260-01